



DECLARATION OF CONFORMITY

This is to certify that this equipment, designed and manufactured by INFICON meets the essential safety requirements of the European Union and is placed on the market accordingly. It has been constructed in accordance with good engineering practice in safety matters in force in the Community and does not endanger the safety of persons, domestic animals or property when properly installed and maintained and used in applications for which it was made.

Equipment Description: D-TEK Select Refrigerant Leak Detector

Model Number: 712-202-Gxx (Applicable to all Group numbers)

Applicable Directives: 2006/95/EC LVD
2004/108/EC General EMC
2011/65/EU RoHS
2006/66/EC as amended by 2013/56/EU Battery Directive

Applicable Standards:

Safety:	EN 61010-1:2010	Safety requirements for electrical equipment for measurement, control, and laboratory use. General requirements.
	IEC 62133:2012	Safety requirements for portable sealed secondary cells, and for batteries made from them, for use in portable applications. CB Test Cert. FI-17925
Emissions:	EN 61326-1:2013	Edition 2.0 (Radiated, Conducted & Harmonic Emissions) (EMC – Measurement, Control & Laboratory Equipment)
	CISPR 11/EN 55011:2009 (+A1:2010)	Emission standard for industrial, scientific, and medical (ISM) radio RF equipment
Immunity:	EN 61326-1:2013	Edition 2.0 (EMC – Measurement, Control & Laboratory Equipment)
		Immunity per Table A.1 – Portable Test and Measurement Equipment
RoHS:	EN 50581:2013	

CE Implementation Date: March 19, 2015


Manufacturer Representative
Brian King
INFICON
Business Line Manager – Service Tools
Two Technology Place
East Syracuse, NY USA 13057

EC Authorized Representative
INFICON GmbH
50968 Köln, Bonner Str. 498

Revision	Description	Date	Approval
--	Initial Issue	03/19/2015	MJV
A	Updated Format per EU requirements Added applicable RoHS standard	11/11/2015	MJV